



Promoting Excellence In Ultrasound

Policies and Statements

A6

Safety Statement On Acoustic Streaming

Policies and Statements

A6

Safety Statement On Acoustic Streaming

*September 1998, **Reaffirmed** February 2000, **Reaffirmed** July 2008*

Flowing movement of particles in liquid-filled anatomical structures and body cavities is commonly seen in diagnostic imaging at higher levels of output used with modern ultrasound equipment.

The radiation force exerted on particles, produces motion that is seen along the axis of the ultrasound beam in the direction away from the transducer.

The streaming velocities involved in clinical diagnostic applications are low, and the limited data on biological effects of streaming suggest that there are no safety concerns.



Australasian Society for Ultrasound in Medicine
PO Box 943, Crows Nest NSW 1585, Sydney AUSTRALIA
Phone: +61 2 9438 2078 • Fax: +61 2 9438 3686
Email: asum@asum.com.au • Website: <http://www.asum.com.au>
ABN: 64 001 679 161